

Notice of Allowability

Application No.

10/059,925

Examiner

Leslie R. Deak

Applicant(s)

HEATH ET AL.

Art Unit

3761

-- The MAILING DATE of this communication appears on the cover sheet with the correspondence address--

All claims being allowable, PROSECUTION ON THE MERITS IS (OR REMAINS) CLOSED in this application. If not included herewith (or previously mailed), a Notice of Allowance (PTOL-85) or other appropriate communication will be mailed in due course. **THIS NOTICE OF ALLOWABILITY IS NOT A GRANT OF PATENT RIGHTS.** This application is subject to withdrawal from issue at the initiative of the Office or upon petition by the applicant. See 37 CFR 1.313 and MPEP 1308.

1. ☒ This communication is responsive to amendment filed 10/16/06.
2. ☒ The allowed claim(s) is/are 1,3-5,7-11,13-25,64,65.
3. ☐ Acknowledgment is made of a claim for foreign priority under 35 U.S.C. § 119(a)-(d) or (f).
 - a) ☐ All b) ☐ Some* c) ☐ None of the:
 1. ☐ Certified copies of the priority documents have been received.
 2. ☐ Certified copies of the priority documents have been received in Application No. _____.
 3. ☐ Copies of the certified copies of the priority documents have been received in this national stage application from the International Bureau (PCT Rule 17.2(a)).

* Certified copies not received: _____.

Applicant has THREE MONTHS FROM THE "MAILING DATE" of this communication to file a reply complying with the requirements noted below. Failure to timely comply will result in ABANDONMENT of this application.

THIS THREE-MONTH PERIOD IS NOT EXTENDABLE.

4. ☐ A SUBSTITUTE OATH OR DECLARATION must be submitted. Note the attached EXAMINER'S AMENDMENT or NOTICE OF INFORMAL PATENT APPLICATION (PTO-152) which gives reason(s) why the oath or declaration is deficient.
 5. ☐ CORRECTED DRAWINGS (as "replacement sheets") must be submitted.
 - (a) ☐ including changes required by the Notice of Draftsperson's Patent Drawing Review (PTO-948) attached
 - 1) ☐ hereto or 2) ☐ to Paper No./Mail Date _____.
 - (b) ☐ including changes required by the attached Examiner's Amendment / Comment or in the Office action of Paper No./Mail Date _____.
- Identifying indicia such as the application number (see 37 CFR 1.84(c)) should be written on the drawings in the front (not the back) of each sheet. Replacement sheet(s) should be labeled as such in the header according to 37 CFR 1.121(d).
6. ☐ DEPOSIT OF and/or INFORMATION about the deposit of BIOLOGICAL MATERIAL must be submitted. Note the attached Examiner's comment regarding REQUIREMENT FOR THE DEPOSIT OF BIOLOGICAL MATERIAL.

Attachment(s)

1. ☐ Notice of References Cited (PTO-892)
2. ☐ Notice of Draftperson's Patent Drawing Review (PTO-948)
3. ☐ Information Disclosure Statements (PTO/SB/08), Paper No./Mail Date _____
4. ☐ Examiner's Comment Regarding Requirement for Deposit of Biological Material
5. ☐ Notice of Informal Patent Application
6. ☐ Interview Summary (PTO-413), Paper No./Mail Date _____
7. ☒ Examiner's Amendment/Comment
8. ☒ Examiner's Statement of Reasons for Allowance
9. ☐ Other _____.

EXAMINER'S AMENDMENT

1. An examiner's amendment to the record appears below. Should the changes and/or additions be unacceptable to applicant, an amendment may be filed as provided by 37 CFR 1.312. To ensure consideration of such an amendment, it MUST be submitted no later than the payment of the issue fee.

The application has been amended as follows:

IN CLAIM 3:

Replace "The catheter of claim 2" in line 1 with --The catheter of claim 1-- .

IN CLAIM 5:

Replace "The catheter of claim 2" in line 1 with --The catheter of claim 1-- .

Reasons for Allowance

2. Claims 1, 3-5, 7-11, 13-25, 64, and 65 are allowed.

3. The following is an examiner's statement of reasons for allowance: The prior art fails to disclose or suggest the apparatus claimed by applicant. In particular, the prior art fails to disclose or suggest a catheter as claimed with proximal and distal openings, a plurality of openings in the sidewall, a septum creating a first and second lumen, wherein the catheter comprises a portion formed in a corkscrew shape with two lumens, one lumen extending distally farther than the other lumen, and one of the openings coincident with the corkscrew portion, along with the other steps and limitations of the claims.

The best prior art of record, US 5,509,897 to Twardowski, discloses the catheter substantially as claimed by applicant with the exception of the corkscrew portion comprising two lumens. Twardowski specifically discloses that one lumen ends before the corkscrew portion of the catheter, and provides no motivation to extend that lumen into the corkscrew portion, rendering the instantly claimed invention patentable over the prior art of record.

Any comments considered necessary by applicant must be submitted no later than the payment of the issue fee and, to avoid processing delays, should preferably accompany the issue fee. Such submissions should be clearly labeled "Comments on Statement of Reasons for Allowance."

Drawings

4. Replacement drawings were received on 22 April 2002. These drawings are accepted.


Conclusion

Any inquiry concerning this communication or earlier communications from the examiner should be directed to Leslie R. Deak whose telephone number is 571-272-4943. The examiner can normally be reached on M-F 7:30-5:00, every other Friday off.

If attempts to reach the examiner by telephone are unsuccessful, the examiner's supervisor, Tanya Zalukaeva can be reached on 571-272-1115. The fax phone number for the organization where this application or proceeding is assigned is 571-273-8300.

Art Unit: 3761

Information regarding the status of an application may be obtained from the Patent Application Information Retrieval (PAIR) system. Status information for published applications may be obtained from either Private PAIR or Public PAIR. Status information for unpublished applications is available through Private PAIR only. For more information about the PAIR system, see <http://pair-direct.uspto.gov>. Should you have questions on access to the Private PAIR system, contact the Electronic Business Center (EBC) at 866-217-9197 (toll-free). If you would like assistance from a USPTO Customer Service Representative or access to the automated information system, call 800-786-9199 (IN USA OR CANADA) or 571-272-1000.



Leslie R. Deak
Patent Examiner
Art Unit 3761
12 December 2006

TATYANA ZALUKAEVA
SUPERVISORY PRIMARY EXAMINER

